

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100367

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/12/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.99: 0.080
LOT #: 07511080A1
EXPIRATION: 03/01/2013

BLANK TEST	0.000	16:00
INTERNAL STANDARD	VERIFIED	16:00
EXTERNAL STANDARD	0.083	16:00
BLANK TEST	0.000	16:01
EXTERNAL STANDARD	0.080	16:01
BLANK TEST	0.000	16:02
EXTERNAL STANDARD	0.081	16:02
BLANK TEST	0.000	16:03
EXTERNAL STANDARD	0.081	16:03
BLANK TEST	0.000	16:04
EXTERNAL STANDARD	0.081	16:05
BLANK TEST	0.000	16:05

Average = 0.0812
Std Dev = 0.0011

Diagnostic Check

VERSIONS

DMT: 1.01
PIC: 3.03
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
Breath Tube = 44.7°C PASSED

PUMP INFO

Flow Rate = 4.455 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9 day of Nov, 2011


Nikki Roth, Notary Public
My Commission Expires With Office




10/13/11